

U.S. DEPARTMENT OF AGRICULTURE  
FOOD SAFETY AND INSPECTION SERVICE  
INTERNATIONAL PROGRAMS

REVIEW DATE

NAME OF FOREIGN LABORATORY

03/15/02

Federal Institute for Veterinary Medicine  
Examinations

## FOREIGN COUNTRY LABORATORY REVIEW

FOREIGN GOV'T AGENCY

Federal Ministry of Social Security and  
Generations

CITY &amp; COUNTRY

Modling, Austria

ADDRESS OF LABORATORY

Robert Kochgasse 17  
2340 Modling, Austria

NAME OF REVIEWER

Dr. Faizur R. Choudry

NAME OF FOREIGN OFFICIAL

Dr. Marina Mikula, Dr. Sepp Flatscher, Deputy Director

Residue Code/Name			100	200	203	400	500	800						
SAMPLING PROCEDURES	REVIEW ITEMS	ITEM #	EVALUATION CODE											
	Sample Handling	01		A	A	A	A	A	A					
	Sampling Frequency	02		A	A	A	A	A	A					
	Timely Analyses	03		C	C	C	C	C	C					
	Compositing Procedure	04		O	O	O	O	O	O					
	Interpret Comp Data	05		O	O	O	O	O	O					
	Data Reporting	06		A	A	A	A	A	A					
ANALYTICAL PROCEDURES	Acceptable Method	07	EVALUATION CODE	A	A	A	A	A	A					
	Correct Tissue(s)	08		A	A	A	A	A	A					
	Equipment Operation	09		C	A	C	C	C	C					
	Instrument Printouts	10		A	A	A	A	A	A					
QUALITY ASSURANCE PROCEDURES	Minimum Detection Levels	11	EVALUATION CODE	A	A	A	A	A	A					
	Recovery Frequency	12		A	A	A	A	A	A					
	Percent Recovery	13		A	A	A	A	A	A					
	Check Sample Frequency	14		A	A	A	A	A	C					
	All analyst w/Check Samples	15		A	A	A	A	A	A					
	Corrective Actions	16		A	A	A	A	A	A					
	International Check Samples	17		A	A	A	A	A	A					
REVIEW PROCEDURES	Corrected Prior Deficiencies	18	EVAL. CODE	O	O	O	O	O	A					
OTHER REVIEW		19	EVAL. CODE											
		20												

SIGNATURE OF REVIEWER

DATE

<b>FOREIGN COUNTRY LABORATORY REVIEW</b> <i>(Comment Sheet)</i>		<b>REVIEW DATE</b> 03/15/02	<b>NAME OF FOREIGN LABORATORY</b> Federal Institute for Veterinary Medicine Examinations
<b>FOREIGN GOV'T AGENCY</b> Federal Ministry of Social Security and Generations	<b>CITY &amp; COUNTRY</b> Modling, Austria		<b>ADDRESS OF LABORATORY</b> Robert Kochgasse 17 2340 Modling, Austria
<b>NAME OF REVIEWER</b> Dr. Faizur R. Choudry		<b>NAME OF FOREIGN OFFICIAL</b> Dr. Marina Mikula, Dr. Sepp Flatscher, Deputy Director	

RESIDUE	ITEM	COMMENTS
100,200, 203,400, 500,800	3	Samples for chlorinated hydrocarbons, trace elements, hormones, chloramphenicol, antibiotics, and sulfonamides were not analyzed in a timely manner such as 80% of samples were analyzed in 42 days. Timely analyses is critical for hormones, antibiotics, and sulfonamides.
100,203, 400,500, 800	9	The standards book for chlorinated hydrocarbons, trace elements, hormones, chloramphenicol, and sulfonamides was not properly maintained for quality assurance program such as: when solutions prepared by the analyst were not signed and verified by the supervisor before the solutions were used; pages were not serially numbered; some standard solutions/reagents/media ingredients, date of purchase and lot number was not recorded.
800	14	The proficiency test (Intralaboratory and/or interlaboratory check samples) for quality assurance program was not performed for sulfonamides.

03/14/02

Institute for Bio-Analytic and Hygiene

**FOREIGN COUNTRY LABORATORY REVIEW**

FOREIGN GOV'T AGENCY  
Private accredited laboratory

CITY & COUNTRY  
Perg, Upper Austria

ADDRESS OF LABORATORY  
Perg, Upper Austria

NAME OF REVIEWER  
Dr. Faiz R. Choudry

NAME OF FOREIGN OFFICIAL  
Dr. Marina Mikula

Residue Code/Name			E.co	Sal														
SAMPLING PROCEDURES	REVIEW ITEMS	ITEM #	EVALUATION CODE															
	Sample Handling	01		A	A													
	Sampling Frequency	02		A	A													
	Timely Analyses	03		A	A													
	Compositing Procedure	04		O	O													
	Interpret Comp Data	05		O	O													
	Data Reporting	06		A	A													
ANALYTICAL PROCEDURES	Acceptable Method	07	EVALUATION CODE	A	A													
	Correct Tissue(s)	08		A	A													
	Equipment Operation	09		A	A													
	Instrument Printouts	10		O	O													
QUALITY ASSURANCE PROCEDURES	Minimum Detection Levels	11	EVALUATION CODE	O	O													
	Recovery Frequency	12		O	O													
	Percent Recovery	13		O	O													
	Check Sample Frequency	14		A	A													
	All analyst w/Check Samples	15		A	A													
	Corrective Actions	16		A	A													
	International Check Samples	17		O	O													
REVIEW PROCEDURES	Corrected Prior Deficiencies	18	EVAL CODE	O	O													
OTHER REVIEW		19	EVAL CODE															
		20																

SIGNATURE OF REVIEWER

DATE

<b>FOREIGN COUNTRY LABORATORY REVIEW</b> <i>(Comment Sheet)</i>		REVIEW DATE 03/14/02	NAME OF FOREIGN LABORATORY Institute for Bio-Analytic and Hygiene
FOREIGN GOV'T AGENCY Private accredited laboratory	CITY & COUNTRY Perg, Upper Austria		ADDRESS OF LABORATORY Perg, Upper Austria
NAME OF REVIEWER Dr. Faiz R. Choudry		NAME OF FOREIGN OFFICIAL Dr. Marina Mikula	

RESIDUE	ITEM	COMMENTS
	14	<p>The proficiency test (Intralaboratory and/or interlaboratory check samples) for quality assurance program was not performed for <i>E.coli</i> and <i>Salmonella</i>.</p> <p>The Institute for Bio-Analytic and Hygiene was accredited by the Ministry of Economic Affairs Accreditation Department in 1997.</p>